



**TATA MEMORIAL CENTRE
HOMI BHABHA CANCER HOSPITAL AND RESEARCH CENTER
MUZAFFARPUR (BIHAR)
(PURCHASE DEPARTMENT)**



CALL FOR OBJECTION

Ref. No. HBCH&RC/MZP/PUR/ST/01/KG

Date: 15.04.2024

Subject: Inviting comments/ objection (if any) regarding procurement of "Qiavac 24 Plus Qiavac Connecting System Vacuum Pump for NGS Lab at HBCH&RC, Muzaffarpur on Proprietary basis

1. Department of NGS Lab, HBCH&RC, Muzaffarpur has requirement of the "**Qiavac 24 Plus Qiavac Connecting System Vacuum Pump**, Proprietary article basis under GFR-166.
2. The proposal submitted by **M/s LCGC Life Sciences LLP, Kolkata** who is sole manufacturer and **M/s Qiagen** is the local agent of these items along with Proprietary Article Certificate attached, same uploaded on TMC & MPMMCC website.
3. The above documents are being uploaded for open information to submit objections/comments (if any) from any manufacturer/ supplier before declaring Proprietary Article of the said equipments / items to be procured. Objections/ comments to be submitted within **05 days** from the date of Issuance/ Uploading of the notification.
4. The objection should be raised in the technical compliance sheet as enclosed, if any firm claiming suitability of their product with respect to specification mentioned.
5. The comments with all relevant & supporting documents should be sent to the Purchase office at MPMMCC, Varanasi in a sealed envelop or through email on purchaseofficer@mpmmcc.tmc.gov.in, purchase@hbchrcmzp.tmc.gov.in within **05** working days from the date of uploading on institutional website i.e. from **16.04.2024** to **20.04.2024**, failing which it will be presumed that any other manufacturer/ vendor is having no comment to offer and case will be decided on merits.


Purchase Officer
HBCH/ MPMMCC, Varanasi

Enclose: Related documents enclosed
1) PAC certification enclosed



Date: 13-03-2024

To Whom it may concern

Sub: Confirmation of the sole manufacturing item of QIAGEN
Ref no: QIPL/AD-1839/Sole source/2023-24

Dear Sir,

Greetings from QIAGEN India Pvt Ltd.

This is with reference to the above subject, we hereby confirm that the below item is proprietary article of QIAGEN, and we are the sole manufacturer of the below mentioned items.

Cat. No	Description of items
19413	QIAvac 24 Plus
19419	QIAvac Connecting System
84020	Vacuum Pump (230 V, 50 Hz)

No company other than QIAGEN, GmbH manufacture the above said items with mentioned Catalog no. Also, QIAGEN India Pvt. Ltd is a direct subsidiary of QIAGEN GmbH and has full-fledged customer & technical support team to provide support in India for the above items. Kindly treat this letter as a confirmation of the proprietary item.

Looking forward for your valuable purchase order.

For QIAGEN India Pvt. Ltd

Arnab Dawn
Account Manager
Authorized Signatory
Contact No. 9205688294



QIAvac

Please quote the model which is including the latest state of art equipment and meets the basic minimum specifications. The quoted model should be latest introduced with end of life not before 12 years from the date of installation at HBCHRC MUZAFFARPUR.

Sr. no.	Description of item	Compliance	
		Yes	No
	Name of the Model:		
	Make:		
	Year of Introduction:		
	Probable end of support:		
1	QIAvac vacuum systems enable fast and efficient vacuum processing of up to 24 spin columns using the QIAvac 24 Plus.		
2	QIAvac vacuum systems are for vacuum purification using QIAGEN spin kits. Vacuum-driven liquid processing facilitates sample handling and reduces hands-on preparation time by eliminating the tedious loading and unloading of spin columns into a centrifuge and subsequent removal of flow-through during binding and washing procedures.		
3	QIAvac 24 Plus is Vacuum manifold for processing 1-24 spin columns: includes QIAvac 24 Plus Vacuum Manifold, Luer Plugs, Quick Couplings		
4	The System comes with Qiavac Connecting system and Vacuum pump (230 V, 50 HZ) for connecting vacuum manifold with vacuum pump: includes Tray, Waste Bottles, Tubing, Couplings, Valve, Gauge, 24 Vac Valves, Universal vacuum pump (capacity 34 L/min, 8 mbar vacuum abs.)		
5	Should perform yearly calibration/Validation and preventive maintenance (2nos) during warranty & if entered into AMC(includes unlimited breakdown calls) /CMC (which will include spare replacement, consumable spares,breakdown calls) Testing & measuring equipment used should be traceable to SI units through National/international standards(As per NABL/NABH norms). To submit SOPs for PMS at the time of installation/commissioning.		
6	Please submit the technical data sheet along with the brochure as a tender document with reference to the quoted make & model.		
7	Please specify list & cost of consumables/consumable Spares(i.e.spares need to be replaced at regular intervals, maybe quarterly/half yearly/yearly such as annual maintenance kit etc.) if any.(Quote Rate)		
8	Please specify pre installation requirements [electrical, HVAC, Compressed air (please specify required air pressure), water requirement (RO/DI/Distilled water with its pressure, flow rate per hour) ; if any of the above is essential]		
9	Demo of the quoted model , will be required if so desired by the user, after the opening of the technical bid and prior to opening of financial bid. This is for technical evaluation.		
10	Predispatch inspection at factory will be required if desired by authorities.		
11	Please visit the site to check whether quoted model can be accomodated in the available space.		
12	Should be CE/FDA/BIS/ICMED approved		
13	Please specify power consumption.		
14	Local Service Support: Should have local office and service support/service engineer for attending the breakdown calls.		
15	Training and Demonstraion is must at the time of installation		
16	Response Time: Should not be more than 12 hrs from lodging a breakdown complaint on toll free or by email.		
17	Warranty : 2 years		
18	AMC : Maximum 2% per annum for 8 years on Ex-works cost after warranty.		
19	Back to back assurance to be taken by the supplier from OEM to supply spares for minimum 10 years and to be submitted.		

- * Only Yes or No should be mentioned in compliance report and additional information should be in remarks column.
- * If 2 years warranty, 8 years AMC/CMC and buyback conditions wherever applicable are not adhered to, TMC reserves right to reject the offer.
- * Please mention whether UPS/Stabilizer is required for the equipment. If so should be included in the offer.

→ Mulkoti B. Am.
 14-3-24
 (265435)

[Signature]
 ASKAR

Barhaddan
 u 168032